

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081423\
 Data File : VD077008.D
 Acq On : 14 Aug 2023 18:32
 Operator : JC/SY
 Sample : 03973-05
 Misc : 5.04G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-03-16.5-17.0

Quant Time: Aug 15 04:50:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080823S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 09 07:25:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	54768	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	114887	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	124694	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	59599	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	38958	90.172	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	180.340%#
35) Dibromofluoromethane	7.804	113	40703	65.900	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	131.800%
50) Toluene-d8	10.263	98	126715	63.874	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	127.740%
62) 4-Bromofluorobenzene	12.575	95	43081	62.120	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	124.240%
Target Compounds						
16) Acetone	4.022	43	3612	27.817	ug/l	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081423\
Data File : VD077008.D
Acq On : 14 Aug 2023 18:32
Operator : JC/SY
Sample : 03973-05
Misc : 5.04G/5.0ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-03-16.5-17.0

Quant Time: Aug 15 04:50:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080823S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 09 07:25:36 2023
Response via : Initial Calibration

