

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081423\
 Data File : VD077000.D
 Acq On : 14 Aug 2023 14:46
 Operator : JC/SY
 Sample : 03971-09
 Misc : 5.05G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B-PP08A

Quant Time: Aug 14 14:57:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080823S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 09 07:25:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	226568	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	457041	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	456362	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	207707	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	93713	52.433	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.860%
35) Dibromofluoromethane	7.805	113	118955	48.412	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.820%
50) Toluene-d8	10.269	98	450931	57.138	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	114.280%
62) 4-Bromofluorobenzene	12.575	95	146491	53.097	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	106.200%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081423\
Data File : VD077000.D
Acq On : 14 Aug 2023 14:46
Operator : JC/SY
Sample : 03971-09
Misc : 5.05G/5.0ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B-PP08A


Quant Time: Aug 14 14:57:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080823S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 09 07:25:36 2023
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD077000.D
Acq: 14 Aug 2023 14:46

Instrument :
MSVOA_D
ClientSampleId :
B-PP08A







Tgt Ion:168 Resp: 226568
Ion Ratio Lower Upper
168 100
99 51.5 36.7 55.1



#33
1,2-Dichloroethane-d4
Concen: 52.433 ug/l
RT: 8.228 min Scan# 1129
Delta R.T. 0.000 min
Lab File: VD077000.D
Acq: 14 Aug 2023 14:46

Tgt Ion: 65 Resp: 93713
Ion Ratio Lower Upper
65 100
67 54.1 0.0 115.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.775 min Scan# 11
Delta R.T. 0.000 min
Lab File: VD077000.D
Acq: 14 Aug 2023 14:46




Instrument :
MSVOA_D
ClientSampleId :
B-PP08A


Tgt Ion:114 Resp: 457041
Ion Ratio Lower Upper
114 100
63 16.4 0.0 33.4
88 13.4 0.0 26.2



#35
Dibromofluoromethane
Concen: 48.412 ug/l
RT: 7.805 min Scan# 1057
Delta R.T. 0.000 min
Lab File: VD077000.D
Acq: 14 Aug 2023 14:46

Tgt Ion:113 Resp: 118955
Ion Ratio Lower Upper
113 100
111 102.2 84.7 127.1
192 18.5 16.8 25.2









#50
Toluene-d8
Concen: 57.138 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD077000.D
Acq: 14 Aug 2023 14:46




Instrument :
MSVOA_D
ClientSampleId :
B-PP08A


Tgt Ion: 98 Resp: 450931
Ion Ratio Lower Upper
98 100
100 66.3 50.3 75.5



#62
4-Bromofluorobenzene
Concen: 53.097 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD077000.D
Acq: 14 Aug 2023 14:46

Tgt Ion: 95 Resp: 146491
Ion Ratio Lower Upper
95 100
174 92.8 0.0 201.8
176 87.2 0.0 191.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 11
Delta R.T. 0.001 min
Lab File: VD077000.D
Acq: 14 Aug 2023 14:46

Instrument :
MSVOA_D
ClientSampleId :
B-PP08A

Tgt Ion:117 Resp: 456362
Ion Ratio Lower Upper
117 100
82 49.4 38.3 57.5
119 32.7 25.0 37.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. -0.000 min
Lab File: VD077000.D
Acq: 14 Aug 2023 14:46

Tgt Ion:152 Resp: 207707
Ion Ratio Lower Upper
152 100
115 57.7 27.1 81.3
150 163.0 0.0 353.6

