

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD081619\
 Data File : VD063633.D
 Acq On : 16 Aug 2019 13:05
 Operator : JC/SY
 Sample : K4356-05RE
 Misc : 2.92g/5ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-3RE

Quant Time: Aug 17 06:46:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 10 06:42:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	846664	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	8.21	114	1184581	50.00	ug/l	-0.03
63) Chlorobenzene-d5	12.36	117	671764	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	14.52	152	198992	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	499055	55.14	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	110.28%	
35) Dibromofluoromethane	6.91	113	541794	54.31	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	108.62%	
50) Toluene-d8	10.39	98	977318	44.56	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	89.12%	
62) 4-Bromofluorobenzene	13.54	95	255965	25.75	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	51.50%	
Target Compounds						
16) Acetone	3.21	43	212852	115.698	ug/l	95
17) Carbon Disulfide	3.36	76	47085	2.181	ug/l	100
20) Methylene Chloride	3.80	84	39812	5.005	ug/l	93
25) 2-Butanone	6.06	43	57228	31.496	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD081619\
Data File : VD063633.D
Acq On : 16 Aug 2019 13:05
Operator : JC/SY
Sample : K4356-05RE
Misc : 2.92g/5ml/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-3RE

Quant Time: Aug 17 06:46:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 10 06:42:47 2019
Response via : Initial Calibration

