

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD081619\
 Data File : VD063643.D
 Acq On : 16 Aug 2019 17:37
 Operator : JC/SY
 Sample : K4389-04
 Misc : 5.95g/5ml/MSVOA D/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 T6-E4(2)

Quant Time: Aug 17 07:13:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 10 06:42:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	1036223	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	8.21	114	1396878	50.00	ug/l	-0.02
63) Chlorobenzene-d5	12.37	117	1008057	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	14.51	152	398258	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	600276	54.19	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	108.38%	
35) Dibromofluoromethane	6.91	113	654623	55.64	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	111.28%	
50) Toluene-d8	10.41	98	1323894	51.18	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	102.36%	
62) 4-Bromofluorobenzene	13.55	95	470904	40.17	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	80.34%	
Target Compounds						
16) Acetone	3.22	43	32746	14.543	ug/l	99
20) Methylene Chloride	3.80	84	93018	9.554	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD081619\
Data File : VD063643.D
Acq On : 16 Aug 2019 17:37
Operator : JC/SY
Sample : K4389-04
Misc : 5.95g/5ml/MSVOA D/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
T6-E4(2)

Quant Time: Aug 17 07:13:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 10 06:42:47 2019
Response via : Initial Calibration

