

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081721\
 Data File : VD070094.D
 Acq On : 17 Aug 2021 18:04
 Operator : VA/SY
 Sample : M3412-02
 Misc : 5.36G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PG-2

Quant Time: Aug 18 03:07:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	486247	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1002437	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	939129	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	334291	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	366149	64.023	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.040%
35) Dibromofluoromethane	7.914	113	351313	54.245	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.500%
50) Toluene-d8	10.337	98	1307468	51.905	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.820%
62) 4-Bromofluorobenzene	12.626	95	376348	45.361	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081721\
Data File : VD070094.D
Acq On : 17 Aug 2021 18:04
Operator : VA/SY
Sample : M3412-02
Misc : 5.36G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PG-2

Quant Time: Aug 18 03:07:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

