

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081721\
 Data File : VD070095.D
 Acq On : 17 Aug 2021 18:32
 Operator : VA/SY
 Sample : M3412-03
 Misc : 6.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PG-3

Quant Time: Aug 18 03:07:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	495712	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1011211	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	964557	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	363999	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	370517	63.550	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.100%
35) Dibromofluoromethane	7.914	113	355264	54.379	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.760%
50) Toluene-d8	10.338	98	1311730	51.622	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.240%
62) 4-Bromofluorobenzene	12.626	95	406823	48.608	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081721\
Data File : VD070095.D
Acq On : 17 Aug 2021 18:32
Operator : VA/SY
Sample : M3412-03
Misc : 6.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PG-3

Quant Time: Aug 18 03:07:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

