

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081722\
 Data File : VD074120.D
 Acq On : 17 Aug 2022 13:17
 Operator : VA/SY
 Sample : VD0817SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0817SBL01

Quant Time: Aug 18 03:17:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081522S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 01:41:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	129385	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	221034	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	207990	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	92803	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	55978	59.556	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.120%
35) Dibromofluoromethane	7.909	113	72917	51.625	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.240%
50) Toluene-d8	10.332	98	173976	40.569	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	81.140%
62) 4-Bromofluorobenzene	12.620	95	81644	46.103	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081722\
Data File : VD074120.D
Acq On : 17 Aug 2022 13:17
Operator : VA/SY
Sample : VD0817SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0817SBL01

Quant Time: Aug 18 03:17:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081522S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 01:41:25 2022
Response via : Initial Calibration

