

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081722\
 Data File : VD074133.D
 Acq On : 17 Aug 2022 19:30
 Operator : VA/SY
 Sample : N4215-03
 Misc : 6.50G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-2

Quant Time: Aug 18 03:23:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081522S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 01:41:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	128910	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	234513	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	222452	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	101129	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	57578	61.773	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.540%
35) Dibromofluoromethane	7.903	113	71937	48.003	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.000%
50) Toluene-d8	10.332	98	161235	35.437	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	70.880%
62) 4-Bromofluorobenzene	12.620	95	79544	42.336	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.680%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081722\
Data File : VD074133.D
Acq On : 17 Aug 2022 19:30
Operator : VA/SY
Sample : N4215-03
Misc : 6.50G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-2

Quant Time: Aug 18 03:23:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081522S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 01:41:25 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.973 min Scan# 1028
Delta R.T. 0.006 min
Lab File: VD074133.D
Acq: 17 Aug 2022 19:30




Instrument :
MSVOA_D
ClientSampleId :
WC-2


Tgt Ion:168 Resp: 128910
Ion Ratio Lower Upper
168 100
99 56.2 49.8 74.6



#33
1,2-Dichloroethane-d4
Concen: 61.773 ug/l
RT: 8.320 min Scan# 1088
Delta R.T. -0.006 min
Lab File: VD074133.D
Acq: 17 Aug 2022 19:30

Tgt Ion: 65 Resp: 57578
Ion Ratio Lower Upper
65 100
67 50.6 0.0 101.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.855 min Scan# 1179

Delta R.T. -0.000 min

Lab File: VD074133.D

Acq: 17 Aug 2022 19:30

Instrument :

MSVOA_D

ClientSampleId :

WC-2





Tgt Ion:114 Resp: 234513

Ion Ratio Lower Upper

114 100

63 21.5 0.0 46.6

88 16.0 0.0 33.2



#35

Dibromofluoromethane

Concen: 48.003 ug/l

RT: 7.903 min Scan# 1017

Delta R.T. -0.006 min

Lab File: VD074133.D

Acq: 17 Aug 2022 19:30




Tgt Ion:113 Resp: 71937


Ion Ratio Lower Upper

113 100

111 101.9 83.8 125.6

192 17.5 14.6 21.8









#50
Toluene-d8
Concen: 35.437 ug/l
RT: 10.332 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VD074133.D
Acq: 17 Aug 2022 19:30




Instrument :
MSVOA_D
ClientSampleId :
WC-2


Tgt Ion: 98 Resp: 161235
Ion Ratio Lower Upper
98 100
100 64.0 52.4 78.6



#62
4-Bromofluorobenzene
Concen: 42.336 ug/l
RT: 12.620 min Scan# 1819
Delta R.T. -0.000 min
Lab File: VD074133.D
Acq: 17 Aug 2022 19:30

Tgt Ion: 95 Resp: 79544
Ion Ratio Lower Upper
95 100
174 81.0 0.0 152.8
176 77.6 0.0 145.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.632 min Scan# 1651
Delta R.T. -0.000 min
Lab File: VD074133.D
Acq: 17 Aug 2022 19:30

Instrument :
MSVOA_D
ClientSampleId :
WC-2

Tgt Ion:117 Resp: 222452
Ion Ratio Lower Upper
117 100
82 53.5 45.3 67.9
119 33.0 26.6 40.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.561 min Scan# 1979
Delta R.T. 0.000 min
Lab File: VD074133.D
Acq: 17 Aug 2022 19:30

Tgt Ion:152 Resp: 101129
Ion Ratio Lower Upper
152 100
115 59.0 31.3 93.8
150 157.7 0.0 348.6

