

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081821\
 Data File : VD070128.D
 Acq On : 18 Aug 2021 15:23
 Operator : VA/SY
 Sample : M3435-05
 Misc : 3.50G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SWCS-21-0002

Quant Time: Aug 19 07:31:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	458420	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	892887	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	772743	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	219797	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	321722	59.669	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.340%
35) Dibromofluoromethane	7.908	113	309477	53.648	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.300%
50) Toluene-d8	10.338	98	1122045	50.009	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.020%
62) 4-Bromofluorobenzene	12.620	95	276201	37.374	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	74.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081821\
Data File : VD070128.D
Acq On : 18 Aug 2021 15:23
Operator : VA/SY
Sample : M3435-05
Misc : 3.50G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SWCS-21-0002

Quant Time: Aug 19 07:31:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

