

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081821\
 Data File : VD070130.D
 Acq On : 18 Aug 2021 16:21
 Operator : VA/SY
 Sample : M3435-09
 Misc : 3.57G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SWCS-22-0306

Quant Time: Aug 19 07:31:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	446196	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	919887	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	773079	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	216638	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	305330	58.181	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 116.360%			
35) Dibromofluoromethane	7.914	113	302292	50.865	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 101.720%			
50) Toluene-d8	10.338	98	1140092	49.322	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 98.640%			
62) 4-Bromofluorobenzene	12.626	95	264160	34.696	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 69.400%			
Target Compounds					Qvalue	
16) Acetone	4.156	43	25537	26.415	ug/l	96
17) Carbon Disulfide	4.403	76	22329	1.165	ug/l	96
52) Toluene	10.396	92	19772	1.132	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081821\
Data File : VD070130.D
Acq On : 18 Aug 2021 16:21
Operator : VA/SY
Sample : M3435-09
Misc : 3.57G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SWCS-22-0306

Quant Time: Aug 19 07:31:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

