

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081821\
 Data File : VD070148.D
 Acq On : 19 Aug 2021 02:51
 Operator : VA/SY
 Sample : M3418-02
 Misc : 6.48G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SWCS-13-1015

Quant Time: Aug 19 07:38:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	518565	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1036033	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1025689	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	474406	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	343703	56.353	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.700%
35) Dibromofluoromethane	7.908	113	344828	51.517	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.040%
50) Toluene-d8	10.338	98	1349781	51.847	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.700%
62) 4-Bromofluorobenzene	12.626	95	467713	54.545	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	109.080%
Target Compounds						
					Qvalue	
16) Acetone	4.156	43	52104	46.374	ug/l	94
17) Carbon Disulfide	4.403	76	47660	2.140	ug/l #	95
25) 2-Butanone	7.208	43	14962	9.745	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081821\
Data File : VD070148.D
Acq On : 19 Aug 2021 02:51
Operator : VA/SY
Sample : M3418-02
Misc : 6.48G/5.00ml/MSVOA_D/SOIL
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SWCS-13-1015

Quant Time: Aug 19 07:38:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

