

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081821\
 Data File : VD070151.D
 Acq On : 19 Aug 2021 04:18
 Operator : VA/SY
 Sample : M3418-05
 Misc : 13.94G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SWCS-14-0515

Quant Time: Aug 19 07:38:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	520605	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	997355	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	960152	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	386604	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	335644	54.816	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.640%	
35) Dibromofluoromethane	7.914	113	334254	51.874	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.740%	
50) Toluene-d8	10.338	98	1273564	50.817	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 101.640%	
62) 4-Bromofluorobenzene	12.626	95	398424	48.266	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 96.540%	
Target Compounds						
					Qvalue	
16) Acetone	4.162	43	107210	95.046	ug/l	92
17) Carbon Disulfide	4.409	76	54075	2.418	ug/l	97
25) 2-Butanone	7.208	43	38211	24.791	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081821\
Data File : VD070151.D
Acq On : 19 Aug 2021 04:18
Operator : VA/SY
Sample : M3418-05
Misc : 13.94G/5.00ml/MSVOA_D/SOIL
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SWCS-14-0515

Quant Time: Aug 19 07:38:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

