


Dichlorodifluoromethane

Response Ratio




$$\text{Response} = 4.711\text{e-}001 * \text{Amt} + 2.554\text{e-}002$$

Coef of Det (r^2) = 0.999817 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D081822S.M

Calibration Table Last Updated: Mon Aug 22 18:40:04 2022


Tert butyl alcohol



Response = $3.463 \times 10^{-2} \times \text{Amt} + 1.061 \times 10^{-2}$
Coef of Det (r^2) = 0.999016 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D081822S.M
Calibration Table Last Updated: Mon Aug 22 18:40:04 2022

Toluene-d8

Response Ratio



$$\text{Response} = 6.074\text{e-}001 * \text{Amt} + 7.727\text{e-}002$$


Coef of Det (r^2) = 0.998228 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D081822S.M

Calibration Table Last Updated: Mon Aug 22 18:40:04 2022


Methacrylonitrile

Response Ratio



$R = -1.140e-002 A^2 + 1.620e-001 A - 1.159e-002$
 Coef of Det (r^2) = 0.992741 Curve Fit: Quadratic
 Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D081822S.M
 Calibration Table Last Updated: Mon Aug 22 18:40:04 2022


Dibromofluoromethane



Response = 2.247e-001 * Amt + 2.034e-002
 Coef of Det (r^2) = 0.999358 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D081822S.M
 Calibration Table Last Updated: Mon Aug 22 18:40:04 2022

1,2-Dichloroethane-d4

Response Ratio



$$\text{Response} = 3.271\text{e-}001 * \text{Amt} + 4.476\text{e-}002$$


Coef of Det (r^2) = 0.998082 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D081822S.M

Calibration Table Last Updated: Mon Aug 22 18:40:04 2022

Bromochloromethane

Response Ratio



$$\text{Response} = 2.995\text{e-}001 * \text{Amt} + 2.189\text{e-}002$$


Coef of Det (r^2) = 0.996198 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D081822S.M

Calibration Table Last Updated: Mon Aug 22 18:40:04 2022

Vinyl Acetate

Response Ratio



$$\text{Response} = 1.018\text{e}+000 * \text{Amt} - 1.542\text{e}-001$$


Coef of Det (r^2) = 0.998173 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D081822S.M

Calibration Table Last Updated: Mon Aug 22 18:40:04 2022

Methylene Chloride

Response Ratio



$$\text{Response} = 5.921\text{e-}001 * \text{Amt} + 1.551\text{e-}001$$


Coef of Det (r^2) = 0.999327 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D081822S.M

Calibration Table Last Updated: Mon Aug 22 18:40:04 2022

Acetone

Response Ratio




$$\text{Response} = 9.406\text{e-}002 * \text{Amt} + 2.089\text{e-}002$$

Coef of Det (r^2) = 0.998765 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D081822S.M

Calibration Table Last Updated: Mon Aug 22 18:40:04 2022

Acrolein



Response = $4.646 \times 10^{-2} \times \text{Amt} - 3.455 \times 10^{-3}$
 Coef of Det (r^2) = 0.999344 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D081822S.M
 Calibration Table Last Updated: Mon Aug 22 18:40:04 2022