

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD081919\
 Data File : VD063656.D
 Acq On : 19 Aug 2019 14:35
 Operator : JC/SY
 Sample : K4406-01
 Misc : 5.03g/5ml/MSVOA D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TT-011-IDWSO-20190816

Quant Time: Aug 20 08:35:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 10 06:42:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.03	168	1131717	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	8.21	114	1457523	50.00	ug/l	-0.03
63) Chlorobenzene-d5	12.36	117	1142756	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	14.52	152	529871	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	525935	43.48	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	86.96%	
35) Dibromofluoromethane	6.92	113	604293	49.23	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	98.46%	
50) Toluene-d8	10.40	98	1271571	47.12	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	94.24%	
62) 4-Bromofluorobenzene	13.55	95	500858	40.95	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	81.90%	
Target Compounds						
16) Acetone	3.22	43	710420	288.892	ug/l	94
20) Methylene Chloride	3.81	84	22667	2.132	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD081919\
Data File : VD063656.D
Acq On : 19 Aug 2019 14:35
Operator : JC/SY
Sample : K4406-01
Misc : 5.03g/5ml/MSVOA D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TT-011-IDWSO-20190816

Quant Time: Aug 20 08:35:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 10 06:42:47 2019
Response via : Initial Calibration

