

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD081919\
 Data File : VD063659.D
 Acq On : 19 Aug 2019 15:56
 Operator : JC/SY
 Sample : K4385-14
 Misc : 4.75µl/5ml/MSVOA D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 T8-C-(0)

Quant Time: Aug 20 08:44:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 10 06:42:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	713230	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	8.22	114	1148772	50.00	ug/l	-0.02
63) Chlorobenzene-d5	12.36	117	559122	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	14.52	152	127452	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	476069	62.44	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	124.88%	
35) Dibromofluoromethane	6.92	113	501739	51.86	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	103.72%	
50) Toluene-d8	10.40	98	724166	34.04	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	68.08%	
62) 4-Bromofluorobenzene	13.55	95	179356	18.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	37.20%	
Target Compounds						
16) Acetone	3.22	43	174939	112.880	ug/l #	92
20) Methylene Chloride	3.81	84	31826	4.749	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD081919\
Data File : VD063659.D
Acq On : 19 Aug 2019 15:56
Operator : JC/SY
Sample : K4385-14
Misc : 4.75g/5ml/MSVOA D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
T8-C-(0)

Quant Time: Aug 20 08:44:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 10 06:42:47 2019
Response via : Initial Calibration

