

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081921\
 Data File : VD070162.D
 Acq On : 19 Aug 2021 12:28
 Operator : VA/SY
 Sample : M3435-05RE
 Misc : 5.62G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SWCS-21-0002RE

Quant Time: Aug 20 05:46:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	490173	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	943533	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	810773	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	251286	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	333354	57.822	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.640%
35) Dibromofluoromethane	7.914	113	319807	52.463	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.920%
50) Toluene-d8	10.337	98	1166441	49.197	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.400%
62) 4-Bromofluorobenzene	12.626	95	302088	38.683	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	77.360%
Target Compounds						
					Qvalue	
16) Acetone	4.155	43	63231	59.537	ug/l	98
25) 2-Butanone	7.208	43	12841	8.848	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081921\
Data File : VD070162.D
Acq On : 19 Aug 2021 12:28
Operator : VA/SY
Sample : M3435-05RE
Misc : 5.62G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SWCS-21-0002RE

Quant Time: Aug 20 05:46:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

