

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081921\
 Data File : VD070173.D
 Acq On : 19 Aug 2021 17:44
 Operator : VA/SY
 Sample : M3418-09
 Misc : 3.22G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SWCS-15-0510

Quant Time: Aug 20 05:48:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	410265	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	785867	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	661021	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	173826	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	312992	64.864	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 129.720%	
35) Dibromofluoromethane	7.914	113	292355	57.582	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 115.160%	
50) Toluene-d8	10.338	98	972396	49.241	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 98.480%	
62) 4-Bromofluorobenzene	12.626	95	221563	34.064	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 68.120%	
Target Compounds						
						Qvalue
16) Acetone	4.150	43	49265	55.422	ug/l	96
17) Carbon Disulfide	4.409	76	21465	1.218	ug/l	97
25) 2-Butanone	7.220	43	13401	11.033	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081921\
Data File : VD070173.D
Acq On : 19 Aug 2021 17:44
Operator : VA/SY
Sample : M3418-09
Misc : 3.22G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SWCS-15-0510

Quant Time: Aug 20 05:48:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

