

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081921\
 Data File : VD070176.D
 Acq On : 19 Aug 2021 19:09
 Operator : VA/SY
 Sample : M3414-13
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B-8

Quant Time: Aug 20 05:48:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	52744	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	92789	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	97452	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	42187	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	47846	77.127	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	154.260%
35) Dibromofluoromethane	7.914	113	33055	55.140	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.280%
50) Toluene-d8	10.332	98	116632	50.021	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.040%
62) 4-Bromofluorobenzene	12.620	95	36721	47.815	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.640%
Target Compounds						
					Qvalue	
16) Acetone	4.156	43	5118	44.785	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081921\
Data File : VD070176.D
Acq On : 19 Aug 2021 19:09
Operator : VA/SY
Sample : M3414-13
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B-8

Quant Time: Aug 20 05:48:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

