

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081921\
 Data File : VD070192.D
 Acq On : 20 Aug 2021 04:43
 Operator : VA/SY
 Sample : M3449-16
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP6

Quant Time: Aug 20 05:51:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	51943	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	91883	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	103670	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	42911	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	52757	86.355	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	172.720%#
35) Dibromofluoromethane	7.914	113	36340	61.217	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	122.440%
50) Toluene-d8	10.337	98	121166	52.478	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.960%
62) 4-Bromofluorobenzene	12.625	95	39969	52.557	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081921\
Data File : VD070192.D
Acq On : 20 Aug 2021 04:43
Operator : VA/SY
Sample : M3449-16
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP6

Quant Time: Aug 20 05:51:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

