

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD082019\
 Data File : VD063682.D
 Acq On : 20 Aug 2019 14:06
 Operator : JC/SY
 Sample : K4385-15RE
 Misc : 5.31g/5ml/MSVOA D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 T8-C-(2)RE

Quant Time: Aug 21 04:07:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 10 06:42:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.03	168	739627	50.00	ug/l	-0.04
34) 1,4-Difluorobenzene	8.20	114	1154229	50.00	ug/l	-0.04
63) Chlorobenzene-d5	12.36	117	583165	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	14.52	152	116715	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	478427	60.51	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	121.02%	
35) Dibromofluoromethane	6.91	113	533197	54.85	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	109.70%	
50) Toluene-d8	10.40	98	771896	36.12	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	72.24%	
62) 4-Bromofluorobenzene	13.55	95	186937m	19.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	38.60%	
Target Compounds						
16) Acetone	3.20	43	154160	95.922	ug/l	97
20) Methylene Chloride	3.79	84	38500	5.540	ug/l #	89
25) 2-Butanone	6.05	43	22710	14.308	ug/l	95
68) m/p-Xylenes	12.67	106	8872	1.329	ug/l	95
73) Isopropylbenzene	13.40	105	12829	1.728	ug/l	87
85) sec-Butylbenzene	14.36	105	15341	2.115	ug/l	93
86) p-Isopropyltoluene	14.49	119	11322	1.632	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD082019\
Data File : VD063682.D
Acq On : 20 Aug 2019 14:06
Operator : JC/SY
Sample : K4385-15RE
Misc : 5.31a/5ml/MSVOA D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
T8-C-(2)RE

Quant Time: Aug 21 04:07:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 10 06:42:47 2019
Response via : Initial Calibration

