

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD082019\
 Data File : VD063686.D
 Acq On : 20 Aug 2019 15:55
 Operator : JC/SY
 Sample : K4388-01
 Misc : 7.95µl/5ml/MSVOA D/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 T4-B3-(0)

Quant Time: Aug 21 03:51:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 10 06:42:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	184	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	0.00	114	0	0.00	ug/l	-8.24
63) Chlorobenzene-d5	12.36	117	239	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	0.00	152	0	0.00	ug/l	-14.53

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	0.00	65	0	0.00	ug/l	
Spiked Amount				50.000		
Recovery				=	0.00%	
35) Dibromofluoromethane	0.00	113	0	0.00	ug/l	
Spiked Amount				50.000		
Recovery				=	0.00%	
50) Toluene-d8	0.00	98	0	0.00	ug/l	
Spiked Amount				50.000		
Recovery				=	0.00%	
62) 4-Bromofluorobenzene	13.65	95	262	0.00	ug/l	0.09
Spiked Amount				50.000		
Recovery				=	0.00%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
6) Chloroethane	2.21	64	193	244.441	ug/l #	42
12) 1,1-Dichloroethene	3.26	96	188	126.447	ug/l #	1
16) Acetone	3.26	43	307	767.854	ug/l #	65
18) Methyl Acetate	3.56	43	214	282.581	ug/l #	57
20) Methylene Chloride	3.84	84	394	227.896	ug/l #	5
22) Diisopropyl ether	5.08	45	188	40.762	ug/l #	33
28) Bromochloromethane	6.59	49	250	215.635	ug/l #	7

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD082019\
Data File : VD063686.D
Acq On : 20 Aug 2019 15:55
Operator : JC/SY
Sample : K4388-01
Misc : 7.95g/5ml/MSVOA D/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
T4-B3-(0)

Quant Time: Aug 21 03:51:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 10 06:42:47 2019
Response via : Initial Calibration

