

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD082019\
 Data File : VD063690.D
 Acq On : 20 Aug 2019 17:45
 Operator : JC/SY
 Sample : K4388-05
 Misc : 6.02g/5ml/MSVOA D/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 T4-B3-(2)

Quant Time: Aug 21 04:37:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 10 06:42:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.03	168	816023	50.00	ug/l	-0.04
34) 1,4-Difluorobenzene	8.21	114	1190656	50.00	ug/l	-0.03
63) Chlorobenzene-d5	12.36	117	817480	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	14.51	152	237622	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	507489	58.18	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	116.36%	
35) Dibromofluoromethane	6.91	113	582141	58.05	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	116.10%	
50) Toluene-d8	10.39	98	1103148	50.04	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	100.08%	
62) 4-Bromofluorobenzene	13.55	95	302202	30.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	60.48%	
Target Compounds						
16) Acetone	3.21	43	21758	12.271	ug/l	98
20) Methylene Chloride	3.80	84	106020	13.828	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD082019\
Data File : VD063690.D
Acq On : 20 Aug 2019 17:45
Operator : JC/SY
Sample : K4388-05
Misc : 6.02g/5ml/MSVOA D/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
T4-B3-(2)

Quant Time: Aug 21 04:37:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 10 06:42:47 2019
Response via : Initial Calibration

