

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD082119\
 Data File : VD063724.D
 Acq On : 21 Aug 2019 20:23
 Operator : JC/SY
 Sample : K4388-12
 Misc : 5.97g/5ml/MSVOA D/SOIL
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 T4-D-(2)

Quant Time: Aug 22 02:12:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 10 06:42:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	988825	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	8.22	114	1310398	50.00	ug/l	-0.02
63) Chlorobenzene-d5	12.37	117	997630	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	376273	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	607745	57.50	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	115.00%	
35) Dibromofluoromethane	6.92	113	690579	62.57	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	125.14%	
50) Toluene-d8	10.41	98	1429723	58.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.84%	
62) 4-Bromofluorobenzene	13.55	95	487921	44.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.74%	
Target Compounds						
16) Acetone	3.22	43	92199	42.911	ug/l	99
20) Methylene Chloride	3.81	84	89692	9.654	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD082119\
Data File : VD063724.D
Acq On : 21 Aug 2019 20:23
Operator : JC/SY
Sample : K4388-12
Misc : 5.97g/5ml/MSVOA D/SOIL
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
T4-D-(2)

Quant Time: Aug 22 02:12:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 10 06:42:47 2019
Response via : Initial Calibration

