

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD082120\
 Data File : VD066369.D
 Acq On : 21 Aug 2020 19:41
 Operator : VA/SY
 Sample : L3734-02
 Misc : 6.74G/5.00ml/MSVOA D/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EROC-BF-1A

Quant Time: Sep 02 07:12:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D082120S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 13:36:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.97	168	198304	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	305246	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	312703	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	163179	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	106861	54.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.94%	
35) Dibromofluoromethane	7.91	113	52078	27.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	54.56%	
50) Toluene-d8	10.34	98	363818	52.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.18%	
62) 4-Bromofluorobenzene	12.63	95	145064	58.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.08%	
Target Compounds						
16) Acetone	4.16	43	28332	80.833	ug/l	98
20) Methylene Chloride	4.96	84	25275	8.201	ug/l #	92
25) 2-Butanone	7.21	43	11317	28.441	ug/l	90
84) 1,2,4-Trimethylbenzene	13.27	105	12857	1.387	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD082120\
Data File : VD066369.D
Acq On : 21 Aug 2020 19:41
Operator : VA/SY
Sample : L3734-02
Misc : 6.74G/5.00ml/MSVOA D/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EROC-BF-1A

Quant Time: Sep 02 07:12:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D082120S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 13:36:12 2020
Response via : Initial Calibration

