


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD082218\
Data File : VD059870.D
Acq On : 22 Aug 2018 11:23
Operator : JC/SY
Sample : BFB
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D082218S.M
Title : SW846 8260
Last Update : Thu Aug 23 03:03:08 2018



AutoFind: Scans 1240, 1241, 1242; Background Corrected with Scan 1235

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.5	41984	PASS
75	95	30	60	41.6	85442	PASS
95	95	100	100	100.0	205210	PASS
96	95	5	9	6.8	14016	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	92.4	189674	PASS
175	174	5	9	7.3	13778	PASS
176	174	95	101	96.9	183808	PASS
177	176	5	9	6.5	12033	PASS