

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD082318\
 Data File : VD059895.D
 Acq On : 23 Aug 2018 18:29
 Operator : JC/SY
 Sample : J4559-08RE
 Misc : 5.01µl/5ml/MSVOA D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RA-081718-SSW-145-052RE

Quant Time: Aug 24 06:50:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D082218S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 03:03:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.40	168	988255	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	7.43	114	1337830	50.00	ug/l	-0.02
63) Chlorobenzene-d5	11.28	117	959774	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	13.37	152	349292	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.77	65	401724	48.23	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	96.46%	
35) Dibromofluoromethane	6.31	113	530339	50.56	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	101.12%	
50) Toluene-d8	9.45	98	1369450	47.05	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	94.10%	
62) 4-Bromofluorobenzene	12.43	95	444178	37.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.14%	
Target Compounds						
20) Methylene Chloride	3.70	84	58094	2.18	ug/l	96
73) Isopropylbenzene	12.27	105	22912	1.03	ug/l	94
85) sec-Butylbenzene	13.21	105	26377	1.11	ug/l	99
86) p-Isopropyltoluene	13.34	119	45477	2.35	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD082318\
Data File : VD059895.D
Acq On : 23 Aug 2018 18:29
Operator : JC/SY
Sample : J4559-08RE
Misc : 5.01a/5ml/MSVOA D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RA-081718-SSW-145-052RE

Quant Time: Aug 24 06:50:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D082218S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 03:03:08 2018
Response via : Initial Calibration

