

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD082318\
 Data File : VD059881.D
 Acq On : 23 Aug 2018 11:01
 Operator : JC/SY
 Sample : VSTDCCC050
 Misc : 5.00g/5ml/MSVOA D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 23 14:48:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D082218S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 03:03:08 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	113	0.00
2 T	Dichlorodifluoromethane	0.447	0.436	2.5	114	0.00
3 P	Chloromethane	0.405	0.414	-2.2	124	0.01
4 C	Vinyl Chloride	0.317	0.343	-8.2#	123	0.00
5 T	Bromomethane	0.055	0.054	1.8	128	0.01
6 T	Chloroethane	0.098	0.098	0.0	110	0.00
7 T	Trichlorofluoromethane	0.367	0.389	-6.0	119	0.00
8 T	Diethyl Ether	0.082	0.083	-1.2	112	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.225	0.239	-6.2	124	0.00
10 T	Methyl Iodide	0.206	0.243	-18.0	108	0.00
11 T	Tert butyl alcohol	0.032	0.030	6.3	98	0.00
12 CM	1,1-Dichloroethene	0.175	0.188	-7.4#	123	0.00
13 T	Acrolein	0.014	0.015	-7.1	108	0.00
14 T	Allyl chloride	0.396	0.413	-4.3	115	0.00
15 T	Acrylonitrile	0.120	0.114	5.0	100	0.00
16 T	Acetone	0.077	0.090	-16.9	123	0.00
17 T	Carbon Disulfide	0.544	0.593	-9.0	120	0.00
18 T	Methyl Acetate	0.160	0.157	1.9	107	0.00
19 T	Methyl tert-butyl Ether	0.991	0.967	2.4	104	0.00
20 T	Methylene Chloride	0.603	0.500	17.1	109	0.00
21 T	trans-1,2-Dichloroethene	0.482	0.474	1.7	109	0.00
22 T	Diisopropyl ether	2.020	2.007	0.6	108	0.00
23 T	Vinyl Acetate	1.129	1.109	1.8	104	0.00
24 P	1,1-Dichloroethane	0.873	0.873	0.0	110	0.00
25 T	2-Butanone	0.227	0.235	-3.5	104	0.00
26 T	2,2-Dichloropropane	0.730	0.728	0.3	117	0.00
27 T	cis-1,2-Dichloroethene	0.528	0.528	0.0	109	0.00
28 T	Bromochloromethane	0.418	0.432	-3.3	111	0.00
29	Tetrahydrofuran	0.108	0.106	1.9	104	0.00
30 C	Chloroform	0.887	0.903	-1.8#	113	0.00
31 T	Cyclohexane	0.802	0.769	4.1	113	0.00
32 T	1,1,1-Trichloroethane	0.753	0.755	-0.3	112	0.00
33 S	1,2-Dichloroethane-d4	0.421	0.421	0.0	109	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	111	0.00
35 S	Dibromofluoromethane	0.392	0.400	-2.0	110	0.00
36 T	1,1-Dichloropropene	0.453	0.454	-0.2	110	0.00
37 T	Ethyl Acetate	0.353	0.341	3.4	106	0.00
38 T	Carbon Tetrachloride	0.457	0.454	0.7	112	0.00
39 T	Methylcyclohexane	0.490	0.496	-1.2	109	0.00
40 TM	Benzene	1.179	1.169	0.8	109	-0.02
41 T	Methacrylonitrile	0.178	0.180	-1.1	104	-0.02
42 TM	1,2-Dichloroethane	0.378	0.385	-1.9	108	0.00
43 T	Isopropyl Acetate	0.459	0.448	2.4	101	0.00
44 TM	Trichloroethene	0.347	0.350	-0.9	110	0.00
45 C	1,2-Dichloropropane	0.329	0.325	1.2#	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD082318\
 Data File : VD059881.D
 Acq On : 23 Aug 2018 11:01
 Operator : JC/SY
 Sample : VSTDCCC050
 Misc : 5.00g/5ml/MSVOA D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 23 14:48:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D082218S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 03:03:08 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.212	0.211	0.5	105	0.00
47 T	Bromodichloromethane	0.480	0.485	-1.0	110	0.00
48 T	Methyl methacrylate	0.267	0.260	2.6	104	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	99	-0.02
50 S	Toluene-d8	1.088	1.115	-2.5	109	0.00
51 T	4-Methyl-2-Pentanone	0.342	0.356	-4.1	103	0.00
52 CM	Toluene	0.756	0.769	-1.7#	112	0.00
53 T	t-1,3-Dichloropropene	0.444	0.450	-1.4	108	0.00
54 T	cis-1,3-Dichloropropene	0.514	0.512	0.4	109	0.00
55 T	1,1,2-Trichloroethane	0.276	0.245	11.2	101	0.00
56 T	Ethyl methacrylate	0.319	0.310	2.8	102	0.00
57 T	1,3-Dichloropropane	0.400	0.392	2.0	103	0.00
58 T	2-Chloroethyl Vinyl ether	0.153	0.151	1.3	111	0.00
59 T	2-Hexanone	0.252	0.258	-2.4	102	0.00
60 T	Dibromochloromethane	0.347	0.347	0.0	106	0.00
61 T	1,2-Dibromoethane	0.279	0.271	2.9	103	0.00
62 S	4-Bromofluorobenzene	0.442	0.437	1.1	107	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	109	0.00
64 T	Tetrachloroethene	0.387	0.386	0.3	108	0.00
65 PM	Chlorobenzene	0.994	1.002	-0.8	107	0.00
66 T	1,1,1,2-Tetrachloroethane	0.363	0.363	0.0	106	0.00
67 C	Ethyl Benzene	1.672	1.697	-1.5#	109	0.00
68 T	m/p-Xylenes	0.609	0.618	-1.5	110	0.00
69 T	o-Xylene	0.595	0.607	-2.0	109	0.00
70 T	Styrene	1.031	1.046	-1.5	109	0.00
71 P	Bromoform	0.314	0.315	-0.3	104	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
73 T	Isopropylbenzene	3.192	3.214	-0.7	111	0.00
74 T	N-amyl acetate	1.398	1.359	2.8	105	0.00
75 P	1,1,2,2-Tetrachloroethane	0.738	0.712	3.5	106	0.00
76 T	1,2,3-Trichloropropane	0.777	0.756	2.7	103	0.00
77 T	Bromobenzene	0.885	0.894	-1.0	109	0.00
78 T	n-propylbenzene	4.042	4.026	0.4	109	0.00
79 T	2-Chlorotoluene	2.286	2.334	-2.1	111	0.00
80 T	1,3,5-Trimethylbenzene	2.538	2.452	3.4	107	0.00
81 T	trans-1,4-Dichloro-2-butene	0.242	0.233	3.7	102	0.00
82 T	4-Chlorotoluene	2.607	2.490	4.5	106	0.00
83 T	tert-Butylbenzene	2.891	2.860	1.1	107	0.00
84 T	1,2,4-Trimethylbenzene	2.645	2.606	1.5	108	0.00
85 T	sec-Butylbenzene	3.399	3.391	0.2	107	0.00
86 T	p-Isopropyltoluene	2.765	2.756	0.3	109	0.00
87 T	1,3-Dichlorobenzene	1.580	1.556	1.5	107	0.00
88 T	1,4-Dichlorobenzene	1.576	1.543	2.1	107	0.00
89 T	n-Butylbenzene	2.782	2.751	1.1	112	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD082318\
Data File : VD059881.D
Acq On : 23 Aug 2018 11:01
Operator : JC/SY
Sample : VSTDCCC050
Misc : 5.00g/5ml/MSVOA D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Aug 23 14:48:25 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D082218S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 03:03:08 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.742	0.734	1.1	107	0.00
91 T	1,2-Dichlorobenzene	1.343	1.300	3.2	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.114	0.112	1.8	104	0.00
93 T	1,2,4-Trichlorobenzene	1.150	1.117	2.9	102	0.00
94 T	Hexachlorobutadiene	0.754	0.778	-3.2	111	0.00
95 T	Naphthalene	1.758	1.733	1.4	103	0.00
96 T	1,2,3-Trichlorobenzene	0.984	0.948	3.7	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6