

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD082319\
 Data File : VD063765.D
 Acq On : 23 Aug 2019 14:44
 Operator : JC/SY
 Sample : K4477-12
 Misc : 6.51g/5ml/MSVOA D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-2(3-4)

Quant Time: Aug 26 01:33:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 10 06:42:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	952539	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	8.21	114	1275829	50.00	ug/l	-0.03
63) Chlorobenzene-d5	12.36	117	1108453	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	14.52	152	631077	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	642430	63.10	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	126.20%	
35) Dibromofluoromethane	6.91	113	686249	63.87	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	127.74%	
50) Toluene-d8	10.40	98	1467076	62.10	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	124.20%	
62) 4-Bromofluorobenzene	13.55	95	621454	58.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.08%	
Target Compounds						
16) Acetone	3.21	43	60380	29.172	ug/l	96
20) Methylene Chloride	3.80	84	91745	10.251	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD082319\
Data File : VD063765.D
Acq On : 23 Aug 2019 14:44
Operator : JC/SY
Sample : K4477-12
Misc : 6.51g/5ml/MSVOA D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-2(3-4)

Quant Time: Aug 26 01:33:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 10 06:42:47 2019
Response via : Initial Calibration

