

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD082319\
 Data File : VD063768.D
 Acq On : 23 Aug 2019 16:05
 Operator : JC/SY
 Sample : K4452-01RE
 Misc : 5.00µl/5ml/MSVOA D/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-CORNER-3RE

Quant Time: Aug 26 11:20:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 10 06:42:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	733069	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	8.21	114	1062775	50.00	ug/l	-0.02
63) Chlorobenzene-d5	12.37	117	733635	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	14.51	152	229687	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	511988	65.34	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	130.68%	
35) Dibromofluoromethane	6.92	113	572821	64.00	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	128.00%	
50) Toluene-d8	10.41	98	1203941	61.18	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	122.36%	
62) 4-Bromofluorobenzene	13.55	95	353437	39.63	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	79.26%	
Target Compounds						
16) Acetone	3.21	43	17072	10.718	ug/l	94
20) Methylene Chloride	3.80	84	16260	2.361	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD082319\
Data File : VD063768.D
Acq On : 23 Aug 2019 16:05
Operator : JC/SY
Sample : K4452-01RE
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-CORNER-3RE

Quant Time: Aug 26 11:20:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 10 06:42:47 2019
Response via : Initial Calibration

