

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082321\
 Data File : VD070228.D
 Acq On : 23 Aug 2021 18:44
 Operator : VA/SY
 Sample : M3510-01
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-02-082021

Quant Time: Aug 24 01:32:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	40619	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	73948	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	77822	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	37933	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	39307	82.276	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 164.560%#		
35) Dibromofluoromethane	7.914	113	27460	57.478	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 114.960%		
50) Toluene-d8	10.338	98	94531	50.873	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 101.740%		
62) 4-Bromofluorobenzene	12.626	95	36471	59.589	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 119.180%		
Target Compounds						
						Qvalue
20) Methylene Chloride	4.956	84	8692	7.575	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082321\
Data File : VD070228.D
Acq On : 23 Aug 2021 18:44
Operator : VA/SY
Sample : M3510-01
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-02-082021

Quant Time: Aug 24 01:32:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

