

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082321\
 Data File : VD070229.D
 Acq On : 23 Aug 2021 19:13
 Operator : VA/SY
 Sample : M3510-03
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-03-082021

Quant Time: Aug 24 01:32:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	25133	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.861	114	48765	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	44222	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	18768	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	25396	85.912	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 171.820%#	
35) Dibromofluoromethane	7.914	113	17095	54.261	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.520%	
50) Toluene-d8	10.332	98	51858	42.320	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 84.640%	
62) 4-Bromofluorobenzene	12.626	95	18304	45.351	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 90.700%	
Target Compounds						Qvalue
20) Methylene Chloride	4.967	84	5594	8.270	ug/l	# 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082321\
Data File : VD070229.D
Acq On : 23 Aug 2021 19:13
Operator : VA/SY
Sample : M3510-03
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-03-082021

Quant Time: Aug 24 01:32:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

