

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082322\
 Data File : VD074202.D
 Acq On : 24 Aug 2022 01:02
 Operator : VA/SY
 Sample : N4260-01
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SASP2-T-4-(0-5)

Quant Time: Aug 24 06:25:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 23 13:12:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	6316	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	16180	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	17954	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	8153	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	8044	124.636	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 249.280%#	
35) Dibromofluoromethane	7.909	113	6943	64.153	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 128.300%	
50) Toluene-d8	10.332	98	19103	54.844	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.680%	
62) 4-Bromofluorobenzene	12.614	95	8077	58.698	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 117.400%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082322\
Data File : VD074202.D
Acq On : 24 Aug 2022 01:02
Operator : VA/SY
Sample : N4260-01
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SASP2-T-4-(0-5)

Quant Time: Aug 24 06:25:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082322S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 23 13:12:17 2022
Response via : Initial Calibration

