

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082322\
 Data File : VD074203.D
 Acq On : 24 Aug 2022 01:31
 Operator : VA/SY
 Sample : N4260-02
 Misc : 5.98G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SASP2-T-4-(5-10)

Quant Time: Aug 24 07:58:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 23 13:12:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	10918	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	25561	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	32136	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	15460	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	14534	130.274	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	260.540%#	
35) Dibromofluoromethane	7.908	113	12839	75.093	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	150.180%#	
50) Toluene-d8	10.332	98	34145	62.052	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	124.100%	
62) 4-Bromofluorobenzene	12.620	95	14784	68.008	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	136.020%	
Target Compounds						
81) trans-1,4-Dichloro-2-b...	12.614	75	7230	123.169	ug/l	Qvalue # 7

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082322\
Data File : VD074203.D
Acq On : 24 Aug 2022 01:31
Operator : VA/SY
Sample : N4260-02
Misc : 5.98G/5.00ml/MSVOA_D/SOIL
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SASP2-T-4-(5-10)

Quant Time: Aug 24 07:58:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082322S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 23 13:12:17 2022
Response via : Initial Calibration

