

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082322\
 Data File : VD074204.D
 Acq On : 24 Aug 2022 02:00
 Operator : VA/SY
 Sample : N4260-03
 Misc : 3.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SASP2-T-4-(10-16)

Quant Time: Aug 24 06:26:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 23 13:12:17 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	80020	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	197662	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	225684	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	106718	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	69591	85.108	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 170.220%#	
35) Dibromofluoromethane	7.908	113	77022	58.256	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 116.520%	
50) Toluene-d8	10.332	98	264610	62.186	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 124.380%	
62) 4-Bromofluorobenzene	12.620	95	95775	56.974	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 113.940%	
Target Compounds						
					Qvalue	
16) Acetone	4.162	43	11077	71.105	ug/l	95
17) Carbon Disulfide	4.409	76	3057	3.165	ug/l #	74
20) Methylene Chloride	4.956	84	16613	12.342	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082322\
Data File : VD074204.D
Acq On : 24 Aug 2022 02:00
Operator : VA/SY
Sample : N4260-03
Misc : 3.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SASP2-T-4-(10-16)

Quant Time: Aug 24 06:26:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082322S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 23 13:12:17 2022
Response via : Initial Calibration

