

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082322\
 Data File : VD074207.D
 Acq On : 24 Aug 2022 03:26
 Operator : VA/SY
 Sample : N4264-01
 Misc : 5.17G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B-3

Quant Time: Aug 24 06:27:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 23 13:12:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	83594	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	215828	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	250876	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	115279	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	77907	91.204	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 182.400%#	
35) Dibromofluoromethane	7.903	113	87076	60.317	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 120.640%	
50) Toluene-d8	10.332	98	291471	62.733	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 125.460%	
62) 4-Bromofluorobenzene	12.620	95	113142	61.640	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 123.280%	
Target Compounds						Qvalue
16) Acetone	4.162	43	2687	16.511	ug/l	92
20) Methylene Chloride	4.967	84	32550	28.854	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082322\
Data File : VD074207.D
Acq On : 24 Aug 2022 03:26
Operator : VA/SY
Sample : N4264-01
Misc : 5.17G/5.00ml/MSVOA_D/SOIL
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B-3

Quant Time: Aug 24 06:27:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082322S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 23 13:12:17 2022
Response via : Initial Calibration

