

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082322\
 Data File : VD074209.D
 Acq On : 24 Aug 2022 04:24
 Operator : VA/SY
 Sample : N4215-05RE
 Misc : 6.96G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-3RE

Quant Time: Aug 24 06:28:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 23 13:12:17 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	75789	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	196649	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	231537	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	108609	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	73943	95.478	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 190.960%#	
35) Dibromofluoromethane	7.908	113	79676	60.574	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 121.140%	
50) Toluene-d8	10.332	98	266070	62.851	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 125.700%	
62) 4-Bromofluorobenzene	12.620	95	106032	63.401	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 126.800%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.961	84	20140	17.622	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082322\
Data File : VD074209.D
Acq On : 24 Aug 2022 04:24
Operator : VA/SY
Sample : N4215-05RE
Misc : 6.96G/5.00ml/MSVOA_D/SOIL
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-3RE

Quant Time: Aug 24 06:28:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082322S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 23 13:12:17 2022
Response via : Initial Calibration

