

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD082418\
 Data File : VD059900.D
 Acq On : 24 Aug 2018 11:46
 Operator : JC/SY
 Sample : VSTDCCC050
 Misc : 5.00g/5ml/MSVOA D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 24 14:50:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D082218S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 03:03:08 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	110	-0.03
2 T	Dichlorodifluoromethane	0.447	0.464	-3.8	118	0.00
3 P	Chloromethane	0.405	0.406	-0.2	119	0.00
4 C	Vinyl Chloride	0.317	0.357	-12.6#	125	0.00
5 T	Bromomethane	0.055	0.059	-7.3	138	0.00
6 T	Chloroethane	0.098	0.106	-8.2	117	0.00
7 T	Trichlorofluoromethane	0.367	0.416	-13.4	125	-0.02
8 T	Diethyl Ether	0.082	0.088	-7.3	116	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.225	0.258	-14.7	131	-0.02
10 T	Methyl Iodide	0.206	0.257	-24.8#	112	-0.02
11 T	Tert butyl alcohol	0.032	0.031	3.1	97	-0.03
12 CM	1,1-Dichloroethene	0.175	0.202	-15.4#	129	-0.02
13 T	Acrolein	0.014	0.013	7.1	94	-0.02
14 T	Allyl chloride	0.396	0.439	-10.9	119	-0.02
15 T	Acrylonitrile	0.120	0.116	3.3	100	-0.03
16 T	Acetone	0.077	0.090	-16.9	120	-0.02
17 T	Carbon Disulfide	0.544	0.607	-11.6	121	-0.02
18 T	Methyl Acetate	0.160	0.161	-0.6	107	-0.02
19 T	Methyl tert-butyl Ether	0.991	0.991	0.0	104	-0.02
20 T	Methylene Chloride	0.603	0.545	9.6	116	-0.02
21 T	trans-1,2-Dichloroethene	0.482	0.504	-4.6	113	-0.03
22 T	Diisopropyl ether	2.020	2.123	-5.1	112	-0.02
23 T	Vinyl Acetate	1.129	1.165	-3.2	107	-0.03
24 P	1,1-Dichloroethane	0.873	0.938	-7.4	116	-0.03
25 T	2-Butanone	0.227	0.240	-5.7	104	-0.02
26 T	2,2-Dichloropropane	0.730	0.783	-7.3	123	-0.03
27 T	cis-1,2-Dichloroethene	0.528	0.572	-8.3	116	-0.03
28 T	Bromochloromethane	0.418	0.457	-9.3	115	-0.03
29	Tetrahydrofuran	0.108	0.108	0.0	103	-0.03
30 C	Chloroform	0.887	0.975	-9.9#	119	-0.03
31 T	Cyclohexane	0.802	0.812	-1.2	116	-0.03
32 T	1,1,1-Trichloroethane	0.753	0.788	-4.6	114	-0.02
33 S	1,2-Dichloroethane-d4	0.421	0.423	-0.5	107	-0.03
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	111	-0.02
35 S	Dibromofluoromethane	0.392	0.405	-3.3	111	-0.03
36 T	1,1-Dichloropropene	0.453	0.480	-6.0	116	-0.03
37 T	Ethyl Acetate	0.353	0.353	0.0	110	-0.03
38 T	Carbon Tetrachloride	0.457	0.474	-3.7	117	-0.02
39 T	Methylcyclohexane	0.490	0.532	-8.6	117	-0.03
40 TM	Benzene	1.179	1.246	-5.7	116	-0.03
41 T	Methacrylonitrile	0.178	0.173	2.8	99	-0.03
42 TM	1,2-Dichloroethane	0.378	0.397	-5.0	111	-0.03
43 T	Isopropyl Acetate	0.459	0.459	0.0	103	-0.03
44 TM	Trichloroethene	0.347	0.359	-3.5	112	-0.03
45 C	1,2-Dichloropropane	0.329	0.349	-6.1#	113	-0.03

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD082418\
 Data File : VD059900.D
 Acq On : 24 Aug 2018 11:46
 Operator : JC/SY
 Sample : VSTDCCC050
 Misc : 5.00µl/5ml/MSVOA D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 24 14:50:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D082218S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 03:03:08 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.212	0.220	-3.8	110	-0.03
47 T	Bromodichloromethane	0.480	0.511	-6.5	115	-0.03
48 T	Methyl methacrylate	0.267	0.270	-1.1	108	-0.02
49 T	1,4-Dioxane	0.002	0.002	0.0	104	-0.03
50 S	Toluene-d8	1.088	1.142	-5.0	111	-0.02
51 T	4-Methyl-2-Pentanone	0.342	0.359	-5.0	104	-0.03
52 CM	Toluene	0.756	0.815	-7.8#	118	-0.03
53 T	t-1,3-Dichloropropene	0.444	0.471	-6.1	113	-0.02
54 T	cis-1,3-Dichloropropene	0.514	0.552	-7.4	117	-0.02
55 T	1,1,2-Trichloroethane	0.276	0.268	2.9	110	-0.03
56 T	Ethyl methacrylate	0.319	0.315	1.3	103	-0.03
57 T	1,3-Dichloropropane	0.400	0.420	-5.0	110	-0.02
58 T	2-Chloroethyl Vinyl ether	0.153	0.151	1.3	110	-0.03
59 T	2-Hexanone	0.252	0.257	-2.0	101	-0.02
60 T	Dibromochloromethane	0.347	0.352	-1.4	107	-0.02
61 T	1,2-Dibromoethane	0.279	0.278	0.4	105	-0.03
62 S	4-Bromofluorobenzene	0.442	0.449	-1.6	110	-0.02
63 I	Chlorobenzene-d5	1.000	1.000	0.0	108	-0.02
64 T	Tetrachloroethene	0.387	0.411	-6.2	115	-0.02
65 PM	Chlorobenzene	0.994	1.064	-7.0	113	-0.02
66 T	1,1,1,2-Tetrachloroethane	0.363	0.378	-4.1	110	0.00
67 C	Ethyl Benzene	1.672	1.761	-5.3#	112	0.00
68 T	m/p-Xylenes	0.609	0.651	-6.9	116	0.00
69 T	o-Xylene	0.595	0.634	-6.6	114	-0.02
70 T	Styrene	1.031	1.079	-4.7	112	-0.02
71 P	Bromoform	0.314	0.319	-1.6	105	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	-0.02
73 T	Isopropylbenzene	3.192	3.379	-5.9	117	0.00
74 T	N-amyl acetate	1.398	1.314	6.0	101	-0.02
75 P	1,1,2,2-Tetrachloroethane	0.738	0.733	0.7	109	-0.02
76 T	1,2,3-Trichloropropane	0.777	0.763	1.8	104	-0.02
77 T	Bromobenzene	0.885	0.965	-9.0	117	0.00
78 T	n-propylbenzene	4.042	4.155	-2.8	112	0.00
79 T	2-Chlorotoluene	2.286	2.379	-4.1	113	0.00
80 T	1,3,5-Trimethylbenzene	2.538	2.699	-6.3	117	-0.02
81 T	trans-1,4-Dichloro-2-butene	0.242	0.235	2.9	103	0.00
82 T	4-Chlorotoluene	2.607	2.677	-2.7	113	-0.02
83 T	tert-Butylbenzene	2.891	3.006	-4.0	112	-0.02
84 T	1,2,4-Trimethylbenzene	2.645	2.688	-1.6	111	0.00
85 T	sec-Butylbenzene	3.399	3.565	-4.9	112	0.00
86 T	p-Isopropyltoluene	2.765	2.874	-3.9	113	-0.02
87 T	1,3-Dichlorobenzene	1.580	1.638	-3.7	113	-0.02
88 T	1,4-Dichlorobenzene	1.576	1.640	-4.1	114	-0.02
89 T	n-Butylbenzene	2.782	2.994	-7.6	122	-0.02

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD082418\
Data File : VD059900.D
Acq On : 24 Aug 2018 11:46
Operator : JC/SY
Sample : VSTDCCC050
Misc : 5.00g/5ml/MSVOA D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Aug 24 14:50:38 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D082218S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 03:03:08 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.742	0.771	-3.9	112	0.00
91 T	1,2-Dichlorobenzene	1.343	1.414	-5.3	113	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.114	0.109	4.4	101	-0.02
93 T	1,2,4-Trichlorobenzene	1.150	1.196	-4.0	109	-0.02
94 T	Hexachlorobutadiene	0.754	0.830	-10.1	118	0.00
95 T	Naphthalene	1.758	1.719	2.2	102	-0.02
96 T	1,2,3-Trichlorobenzene	0.984	1.023	-4.0	109	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 6