

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD082420\
 Data File : VD066379.D
 Acq On : 24 Aug 2020 14:19
 Operator : VA/SY
 Sample : L3734-07
 Misc : 7.23G/5.00ml/MSVOA D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EROC-BF-1F

Manual Integrations
 APPROVED

MMDadoda
 8/25/2020 3:32:57 PM

Quant Time: Aug 25 03:42:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D082120S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 13:36:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.97	168	191237	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	292002	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	275713	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	145020	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	105066	56.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.08%	
35) Dibromofluoromethane	7.91	113	56726	31.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	62.14%	
50) Toluene-d8	10.34	98	342131	51.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.42%	
62) 4-Bromofluorobenzene	12.63	95	125200	52.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.62%	
Target Compounds						
8) Diethyl Ether	3.71	74	1571m	2.084	ug/l	Qvalue
16) Acetone	4.16	43	10121	29.943	ug/l #	72
20) Methylene Chloride	4.96	84	26049	9.262	ug/l	93
44) Trichloroethene	9.11	130	7857	3.450	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD082420\
Data File : VD066379.D
Acq On : 24 Aug 2020 14:19
Operator : VA/SY
Sample : L3734-07
Misc : 7.23G/5.00ml/MSVOA D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EROC-BF-1F

Manual Integrations
APPROVED

MMDadoda
8/25/2020 3:32:57 PM

Quant Time: Aug 25 03:42:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D082120S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 13:36:12 2020
Response via : Initial Calibration

