

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082521\
 Data File : VD070242.D
 Acq On : 25 Aug 2021 10:46
 Operator : VA/SY
 Sample : VD0825SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0825SBL01

Quant Time: Aug 25 14:43:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	514929	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1041322	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1027702	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	446041	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	349160	57.652	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.300%
35) Dibromofluoromethane	7.914	113	351045	52.180	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.360%
50) Toluene-d8	10.338	98	1376236	52.595	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.180%
62) 4-Bromofluorobenzene	12.626	95	445809	51.726	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082521\
Data File : VD070242.D
Acq On : 25 Aug 2021 10:46
Operator : VA/SY
Sample : VD0825SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0825SBL01

Quant Time: Aug 25 14:43:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

