

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082521\
 Data File : VD070248.D
 Acq On : 25 Aug 2021 13:45
 Operator : VA/SY
 Sample : M3465-03
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CON-B

Quant Time: Aug 25 14:46:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	408722	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	813931	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	822573	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	354063	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	277722	57.772	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.540%
35) Dibromofluoromethane	7.908	113	166842	31.728	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	63.460%
50) Toluene-d8	10.338	98	1080106	52.810	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.620%
62) 4-Bromofluorobenzene	12.626	95	351720	52.210	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.420%
Target Compounds						
16) Acetone	4.162	43	14729	16.632	ug/l	90
20) Methylene Chloride	4.962	84	64503	3.030	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082521\
Data File : VD070248.D
Acq On : 25 Aug 2021 13:45
Operator : VA/SY
Sample : M3465-03
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CON-B

Quant Time: Aug 25 14:46:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

