

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082521\
 Data File : VD070253.D
 Acq On : 25 Aug 2021 16:09
 Operator : VA/SY
 Sample : M3524-02
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PE-2

Quant Time: Aug 26 02:45:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	24300	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.861	114	36710	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	38857	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	16484	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	25578	89.494	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 178.980%#	
35) Dibromofluoromethane	7.914	113	14755	62.213	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 124.420%	
50) Toluene-d8	10.338	98	45450	49.270	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 98.540%	
62) 4-Bromofluorobenzene	12.626	95	16118	53.049	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 106.100%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082521\
Data File : VD070253.D
Acq On : 25 Aug 2021 16:09
Operator : VA/SY
Sample : M3524-02
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PE-2

Quant Time: Aug 26 02:45:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

