

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082823\
 Data File : VD077044.D
 Acq On : 28 Aug 2023 16:32
 Operator : JC/SY
 Sample : 04183-03
 Misc : 5.02G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HR-03-082523

Quant Time: Aug 29 02:51:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082823S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 29 02:15:23 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	60292	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	132750	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	139641	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	68464	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	47455	79.943	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	159.880%	
35) Dibromofluoromethane	7.804	113	49926	56.295	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	112.600%	
50) Toluene-d8	10.269	98	159348	49.666	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	99.340%	
62) 4-Bromofluorobenzene	12.569	95	53221	56.262	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	112.520%	
Target Compounds						
20) Methylene Chloride	4.799	84	11178	10.060	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082823\
Data File : VD077044.D
Acq On : 28 Aug 2023 16:32
Operator : JC/SY
Sample : 04183-03
Misc : 5.02G/5.0ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HR-03-082523

Quant Time: Aug 29 02:51:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082823S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 29 02:15:23 2023
Response via : Initial Calibration

