

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082823\
 Data File : VD077047.D
 Acq On : 28 Aug 2023 17:56
 Operator : JC/SY
 Sample : 04165-01
 Misc : 5.06G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VNJ-208

Quant Time: Aug 29 02:51:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082823S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 29 02:15:23 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	17371	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	40115	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	45503	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	21295	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	17970	105.070	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 210.140%#		
35) Dibromofluoromethane	7.799	113	16177	60.363	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 120.720%		
50) Toluene-d8	10.263	98	48684	50.214	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	= 100.420%		
62) 4-Bromofluorobenzene	12.575	95	16960	59.332	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	= 118.660%		
Target Compounds						
20) Methylene Chloride	4.787	84	4204	14.415	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082823\
Data File : VD077047.D
Acq On : 28 Aug 2023 17:56
Operator : JC/SY
Sample : 04165-01
Misc : 5.06G/5.0ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VNJ-208

Quant Time: Aug 29 02:51:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082823S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 29 02:15:23 2023
Response via : Initial Calibration

