

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082823\
 Data File : VD077048.D
 Acq On : 28 Aug 2023 18:24
 Operator : JC/SY
 Sample : 04169-01
 Misc : 5.06G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PL-01-082423

Quant Time: Aug 29 02:52:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082823S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 29 02:15:23 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	31593	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	68673	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	76386	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	37032	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	28572	91.856	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 183.720%#		
35) Dibromofluoromethane	7.804	113	29090	63.407	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 126.820%		
50) Toluene-d8	10.269	98	82630	49.785	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	= 99.560%		
62) 4-Bromofluorobenzene	12.575	95	29373	60.025	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	= 120.040%		
Target Compounds						
20) Methylene Chloride	4.787	84	6937	12.689	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082823\
Data File : VD077048.D
Acq On : 28 Aug 2023 18:24
Operator : JC/SY
Sample : 04169-01
Misc : 5.06G/5.0ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PL-01-082423

Quant Time: Aug 29 02:52:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082823S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 29 02:15:23 2023
Response via : Initial Calibration

