

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082923\
 Data File : VD077060.D
 Acq On : 29 Aug 2023 14:23
 Operator : JC/SY
 Sample : 04196-17
 Misc : 3.10G/5.1ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-345-VOC-03

Quant Time: Aug 29 21:00:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082823S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 29 02:15:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	119358	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	274936	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.580	117	297591	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	128004	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	72865	62.005	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 124.000%	
35) Dibromofluoromethane	7.804	113	98462	53.607	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.220%	
50) Toluene-d8	10.269	98	357371	53.781	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 107.560%	
62) 4-Bromofluorobenzene	12.574	95	102164	52.148	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 104.300%	
Target Compounds						
20) Methylene Chloride	4.798	84	36757	19.487	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082923\
Data File : VD077060.D
Acq On : 29 Aug 2023 14:23
Operator : JC/SY
Sample : 04196-17
Misc : 3.10G/5.1ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-345-VOC-03

Quant Time: Aug 29 21:00:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082823S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 29 02:15:23 2023
Response via : Initial Calibration

