

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082923\
 Data File : VD077061.D
 Acq On : 29 Aug 2023 14:52
 Operator : JC/SY
 Sample : 04196-24
 Misc : 3.22G/5.2ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-345-VOC-04

Quant Time: Aug 29 21:00:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082823S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 29 02:15:23 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	116120	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	268391	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	297494	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	135772	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	70485	61.652	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 123.300%	
35) Dibromofluoromethane	7.799	113	96829	54.003	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.000%	
50) Toluene-d8	10.269	98	351859	54.243	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 108.480%	
62) 4-Bromofluorobenzene	12.575	95	108077	56.511	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 113.020%	
Target Compounds						
20) Methylene Chloride	4.805	84	33598	18.055	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082923\
Data File : VD077061.D
Acq On : 29 Aug 2023 14:52
Operator : JC/SY
Sample : 04196-24
Misc : 3.22G/5.2ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-345-VOC-04

Quant Time: Aug 29 21:00:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082823S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 29 02:15:23 2023
Response via : Initial Calibration

