

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082923\
 Data File : VD077062.D
 Acq On : 29 Aug 2023 15:20
 Operator : JC/SY
 Sample : 04196-31
 Misc : 5.20G/5.3ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-345-VOC-05

Quant Time: Aug 29 21:00:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082823S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 29 02:15:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	116388	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	275094	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.580	117	297895	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	137954	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	70419	61.452	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.900%
35) Dibromofluoromethane	7.804	113	94642	51.497	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.000%
50) Toluene-d8	10.269	98	355696	53.498	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	107.000%
62) 4-Bromofluorobenzene	12.575	95	109361	55.789	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	111.580%
Target Compounds						
20) Methylene Chloride	4.793	84	36732	20.075	ug/l	Qvalue # 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082923\
Data File : VD077062.D
Acq On : 29 Aug 2023 15:20
Operator : JC/SY
Sample : 04196-31
Misc : 5.20G/5.3ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-345-VOC-05

Quant Time: Aug 29 21:00:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082823S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 29 02:15:23 2023
Response via : Initial Calibration

