

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082923\
 Data File : VD077067.D
 Acq On : 29 Aug 2023 17:41
 Operator : JC/SY
 Sample : 04198-03
 Misc : 4.52G/5.6ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-309-OK-370-VOC-24

Quant Time: Aug 29 21:01:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082823S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 29 02:15:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	116199	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	264758	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	302285	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	140613	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	73987	64.671	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 129.340%	
35) Dibromofluoromethane	7.805	113	94438	53.392	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.780%	
50) Toluene-d8	10.269	98	351913	54.996	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 110.000%	
62) 4-Bromofluorobenzene	12.575	95	113192	59.998	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 120.000%	
Target Compounds						
20) Methylene Chloride	4.793	84	29269	15.174	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082923\
Data File : VD077067.D
Acq On : 29 Aug 2023 17:41
Operator : JC/SY
Sample : 04198-03
Misc : 4.52G/5.6ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-309-OK-370-VOC-24

Quant Time: Aug 29 21:01:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082823S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 29 02:15:23 2023
Response via : Initial Calibration

