

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082923\
 Data File : VD077068.D
 Acq On : 29 Aug 2023 18:09
 Operator : JC/SY
 Sample : 04198-09
 Misc : 6.24G/5.7ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-309-OK-370-VOC-25

Quant Time: Aug 29 21:01:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082823S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 29 02:15:23 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	115976	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	260590	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	297462	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	138622	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	70699	61.916	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 123.840%	
35) Dibromofluoromethane	7.804	113	93241	53.559	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.120%	
50) Toluene-d8	10.269	98	346609	55.033	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 110.060%	
62) 4-Bromofluorobenzene	12.575	95	106914	57.577	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 115.160%	
Target Compounds						
20) Methylene Chloride	4.793	84	28969	15.013	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082923\
Data File : VD077068.D
Acq On : 29 Aug 2023 18:09
Operator : JC/SY
Sample : 04198-09
Misc : 6.24G/5.7ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-309-OK-370-VOC-25

Quant Time: Aug 29 21:01:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082823S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 29 02:15:23 2023
Response via : Initial Calibration

