

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082923\
 Data File : VD077069.D
 Acq On : 29 Aug 2023 18:38
 Operator : JC/SY
 Sample : 04198-16
 Misc : 1.42G/5.8ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-309-OK-370-VOC-26

Quant Time: Aug 29 21:01:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082823S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 29 02:15:23 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	116496	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	262085	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	288587	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	131552	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	69733	60.797	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	121.600%	
35) Dibromofluoromethane	7.804	113	93821	53.585	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	107.160%	
50) Toluene-d8	10.269	98	342772	54.114	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	108.220%	
62) 4-Bromofluorobenzene	12.575	95	106093	56.808	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	113.620%	
Target Compounds						
20) Methylene Chloride	4.799	84	17325	7.239	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082923\
Data File : VD077069.D
Acq On : 29 Aug 2023 18:38
Operator : JC/SY
Sample : 04198-16
Misc : 1.42G/5.8ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-309-OK-370-VOC-26

Quant Time: Aug 29 21:01:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082823S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 29 02:15:23 2023
Response via : Initial Calibration

